

Figure: 25 TAC §289.230(x)(3)

Specific Subsection	Name of Record	Time Interval for Record Keeping
(h)(1)(A)	Interpreting Physician Qualifications	Until 2 years after terminating certification or 2 years after the physician becomes inactive at the facility
(h)(1)(C)	Interpreting Physician Continuing Education and Experience	6 years
(h)(1)(E)	Mandatory training for Interpreting Physician, if applicable	6 years
(h)(2)(A)	Medical Radiologic Technologist (MRT) Qualifications	Until 2 years after terminating certification or 2 years after the MRT becomes inactive at the facility.
(h)(2)(C)	Medical Radiologic Technologist Continuing Education and Experience	6 years
(h)(2)(E)	Mandatory training for Medical Radiologic Technologist, if applicable	6 years
(h)(3)(A)	Medical Physicist Qualifications	Until 2 years terminating certification or 2 years after the physicist becomes inactive at the facility
(h)(3)(C)	Medical Physicist Continuing Education and Experience	6 years
(i)(10)	FDA Variances	Until termination of certification or equipment is replaced
(k)(2)	Quality Assurance (QA) Records	Until the next annual inspection has been completed and the department has determined that the facility is compliant with the QA requirements or until the test has been performed two additional times at the required frequency, whichever is longer.
(l)(10)	Physicist Mammography Survey	7 years
(l)(11)	Physicist Mammography Equipment Evaluation	2 years

(m)(2)	Medical Outcomes Audit	2 years
(o)	Complaints	3 years
(t)(1)	Operating & Safety Procedures	Until termination of certification
(t)(5); (w)(13)	Records of Receipt, Transfer, and Disposal	Until termination of certification
(t)(8)(B)	Protective Devices Annual Check	3 years
(t)(10)	Records on Calibration, Maintenance and Modifications Performed on Mammography Machines	2 years
(t)(1)(A)	Current §§289.203, 289.204, 289.205, 289.226, 289.227, 289.230, and 289.231.	Until termination of certification
(k)(2)	Current Certification of Mammography Systems	Until termination of certification
(f)(2)	Current Accreditation of Mammography Systems	Until termination of certification
(v)(5)	Certification of Inspection	Until termination of certification
(v)(6)	Notice of Failure	Until termination of certification
(v)(7)	Patient Notification	Until termination of certification
(w)(14)	Records of Calibration, Maintenance, and Modifications Performed on Interventional Breast Radiography Machines	Until termination of certification