Figure: 25 TAC §289.229(h)(5)(F)(viii)

The average ratios of exposure in milliRoentgen (mR) to the indicated mAs product obtained at any two consecutive Milliampere (mA) or Milliampere-seconds (mAs) settings must not differ by more than 0.10 times their sum, where X_1 and X_2 are the average mR/mAs values obtained at each of two consecutive tube current settings:

 $|x\bar{x}_1 - x\bar{x}_2| \le 0.10(x\bar{x}_1 + x\bar{x}_2)$