Figure: 25 TAC §289.229(I)

	Name of Record	Rule Cross-Reference	Time Interval Required forRecord Keeping	
Accelerators used for research and development and industrial operations				
(A)	Initial surveys	(f)(2)(C)	Until termination of registration	
(B)	Tests and repairs	(f)(3)(A)(x)	5 years	
(C)	Calibration, surveys	(f)(3)(F)	5 years	
(D)	Contamination smear for units operating greater than 10 MeV	(f)(3)(G)	Until termination of registration	
(E)	Receipt, transfers, and disposal	(f)(3)(H)	Until termination of registration	
(F)	Training for operators	(f)(4)(B)	Until 2 years after the individual terminates employment	
Therapeutic radiation machines, radiation therapy simulation systems, and EBT devices				
(G)	Credentials of operators	(h)(1)(C)	Until 2 years after the individual	
	EBT device operators	(h)(1)(E)	terminates employment	
(H)	Review of qualityassurance program	(h)(1)(F)(vii)	5 years	
(I)	Written OSP	(h)(1)(G)	Until transfer of machine or termination of registration	
(L)	FDA variances	(h)(1)(H)	Until transfer of machine or termination of registration	
(K)	Initial and SubsequentSurveys		Until termination of registration	
	Therapy (below 1 MeV)	(h)(2)(D)(i)(II)		
	Therapy (1 MeV and above)	(h)(3)(C)(i)(III)		
	EBT device	(h)(4)(B)(i)(II)		

(L)	Calibration		5 years
	Therapy (below 1 MeV)	(h)(2)(D)(ii)(V)	
	Therapy (1 MeV and above)	(h)(3)(C)(ii)	
	EBT device	(h)(4)(B)(ii)	
(M)	Contamination smears for units operating greater than 10 MeV	(h)(1)(I)	Until termination of registration
(N)	QA checks and corrective actions		5 years after the QA checks
	Therapy (below 1 MeV)	(h)(2)(D)(iii)(VI)	
	Therapy (1 MeV and above)	(h)(3)(C)(iii)(VII)	
	EBT device	(h)(4)(B)(iii)	
(0)	Leakage measurements		5 years
	Therapy (1 MeV and above)	(h)(3)(A)(i)	
(P)	Protective devices forradiation therapy simulation systems	(h)(5)(A)(iii)(II)	3 years
(Q)	Film processing recordsfor simulators	(h)(5)(E)(i)(V), (VI), and (ii)	3 years
(R)	Digital imaging acquisition systems	(h)(5)(E)(iii)	3 years
(S)	CT dose measurements	(h)(5)(C)(iv)(III)	5 years
(T)	CT films resulting from quality control tests	(h)(5)(D)(ii)	1 year or until a new phantomimage is performed
(U)	Reports of medical events	(j)(2)	Until termination of registration